

PREMARKET NOTIFICATION

JUN 19 2013

510(k) SUMMARY

(As Required By 21 CFR 807.92)

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and 21 CFR §807.92.

The assigned 510(k) number is: _____ Date: March 01 2013

1. Submitter:

Health & Life Co., Ltd.

9F, No.186, Jian Yi Road, Zhonghe District, New Taipei City, Taiwan, R.O.C.

TEL: +886-2-8227-1300

FAX: +886-2-8227-1301

Contact person: Sarah Su/ Regulatory Affairs Dept.

E-mail: sarah.su@hlmt.com.tw

Tel: 886-2-8227-1300 ext.1201

Fax: 886-2-8227-1301

2. Name of the Device:

Trade Name: Full Automatic (NIBP) Blood Pressure Monitor, Model HL158LA

Common Name: Blood Pressure Monitor

Classification Name: Non-invasive Blood Pressure Measurement System

Classification: Class II, 21CFR 870.1130

Product Code: DXN

Panel: 74 Cardiovascular



3. Information for the 510(k) Cleared Device (Predicate Device):

A. Full Automatic (NIBP) Blood Pressure Monitor, Model HL158BT, K093824

B. A&D Digital Blood Pressure Monitor, Model UB-512, K042967

4. Device Description:

HL158LA uses the oscillometric method to automatically measure systolic and diastolic blood pressure as well as heart rate. The measurement position is at human being's wrist. All values can be read out in one LCD panel. The device is designed for home use and recommended for use by adults aged 18 years and older with wrist circumference ranging approx. 5.3 ~ 7.7 inch (13.5 ~ 19.5 cm).

The device is equipped with Irregular Heartbeat Detector function. It will display a symbol of  or , to indicate the detection of irregular heartbeat rhythm as defined as a rhythm is more than or less than 25% from the average heartbeat during the measurement.

Besides, user can save and manage the measurement data by transferring the measured readings of blood pressure to the connected personal computer (PC) via USB cable. Additionally, the Risk Category Indicator (Six Levels) feature will classify the blood pressure results with WHO (World Health Organization) Classifications, which are Sever Hypertension, Moderate Hypertension, Mild Hypertension, High Normal, Normal, and Optimal. The Corresponding LCD segment will be turned on along with the systolic, diastolic, and heart rate information.

5. Intended Use

HL158LA automatically measures human's Systolic, Diastolic blood pressure and heart rate by using the oscillometric method. All values can be read out in one LCD panel. Measurement position is at human being's wrist. The intended use of this over-the-counter device is for use by people over the age of 18 with wrist circumference ranging from approx. 5.3 to 7.7 inches(13.5 cm to 19.5 cm) and for home use.

6. Comparison of device to predicate device:

Product Specification Comparison Table of HL158LA and HL158BT (K093824)

Item	Predicate Device HL158BT (K093824)	Subject Device HL158LA
Method of measurement	Oscillometric	Same as left
Measurement range of	Pressure 0- 300mmHg Pulse 40-199 beats/minute	Rated range of cuff pressure: 0- 300 mmHg Rated range of determination: 40- 280mmHg Pulse: 40-199 beats/minute
Accuracy	Pressure \pm 3 mmHg Pulse \pm 5%	Same as left
Inflation	Automatic inflation (Air pump)	Same as left
Deflation of pressure	Automatic air release control valve	Same as left
Display	Liquid Crystal Display	Same as left

Memory	60 memory * 2 users (120 memory total)	40 memory * 3 users (120 memory total)
Cuff size	Wrist circumference approx. 5.3~7.7 inch (13.5 ~ 19.5 cm)	Same as left
Operation environment	50°F ~ 104°F (10°C ~ 40°C), 15%~90%R.H.	41°F~104°F (5°C ~ 40°C), 15%~93% R.H.
Storage/ Transportation environment	-4°F ~ +158°F (-20°C ~ +70°C), ≤ 90 % R.H.	- 13°F~+158°F (-25°C ~ +70°C), ≤ 93%R.H.
Power Supply	2 × “AAA” (1.5V) Alkaline battery	Same as left
Material	ABS housing and rubber keys	ABS housing and ABS keys
Number of Push Button	7	4
Storage case	Yes	Same as left
Unit Weight	Approx. 106 ± 5 g (Excluding batteries)	Approx. 119 ± 5 g (Excluding cuff and batteries)

Changes from the predicate devices HL158BT (K093824):

* 4 push buttons, changing of exterior casing design and sets of memory

* Additional modified feature of Risk Category Indicator (Six Levels)

For the modified feature of Risk Category Indicator (Six Levels), was compared with the other predicate device A&D UB-512 (K042967).

7. Discussion of Clinical Tests Performed:

HL158LA is compliant to the ANSI/AAMI SP10:2002/(R)2008 & ANSI/AAMI SP10:2002/A1:2003/(R)2008 & ANSI/AAMI SP10:2002/A2:2006/(R)2008 Standard for Manual, electronic, or automated sphygmomanometers. All the relevant activities were performed by designate individual(s) and the results demonstrated that the predetermined acceptance criteria were fully met.

8. Discussions of Non-Clinical Tests Performed for Determination of Substantial Equivalence are as follows:

The subject device was tested to evaluate its safety and effectiveness, including the followings:

a. Safety Test

- IEC 60601-1:2005, Medical electrical equipment - Part 1: General requirements for basic safety and essential performance.

- IEC 60601-1-11:2010, Medical electrical equipment - Part 1-11: General requirements for basic safety and essential performance - Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment.
- b. **EMC Test:** IEC 60601-1-2 Edition 3:2007-03, Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests.
- c. **FCC Test:** FCC 47 CFR Part 15, Subpart B:2011
- d. **Biocompatibility Test:**
 - ISO 10993-1:2009, Biological evaluation of medical devices -- Part 1: Evaluation and testing within a risk management process.
 - ISO 10993-5:2009, Biological evaluation of medical devices -- Part 5: Tests for In Vitro cytotoxicity.
 - ISO 10993-10:2010 Third Edition, Biological evaluation of medical devices -- Part 10: Tests for irritation and skin sensitization
- e. **Reliability Test:** ANSI/AAMI SP10:2002/(R)2008 & ANSI/AAMI SP10:2002/A1:2003/(R)2008 & ANSI/AAMI SP10:2002/A2:2006/(R)2008 Manual, electronic or automated sphygmomanometers.
- f. **Risk Assessment:** ISO 14971:2007 Second Edition, Medical devices – Application of risk management to medical devices.

9. Conclusions:

The subject device was tested and fulfilled the requirements from those standards mentioned above, and it's concluded that the subject device is substantially equivalent to the predicate devices.



June 19, 2013

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-002

Health & Life Co., Ltd.
Sarah Su
9f, No.186, Jian Yi Road
Zhonghe District, New Taipei City, 23553 TW

Re: K130564
Trade/Device Name: Full automatic (nibp) blood pressure monitor
Regulation Number: 21 CFR 870.1130
Regulation Name: Blood Pressure Monitor
Regulatory Class: Class II
Product Code: DXN
Dated: May 14, 2013
Received: May 20, 2013

Dear Sarah Su:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, Misbranding by reference to premarket notification (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Owen P. Faris -S

for

Bram D. Zuckerman, Ph.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indication for Use

510(k) Number (if known): _____

Device Name: Full Automatic (NIBP) Blood Pressure Monitor, Model HL158LA**Indications for Use:**

HL158LA automatically measures human's Systolic, Diastolic blood pressure and heart rate by using the oscillometric method. All values can be read out in one LCD panel. Measurement position is at human being's wrist. The intended use of this over-the-counter device is for use by people over the age of 18 with wrist circumference ranging from approx. 5.3 to 7.7 inches(13.5 cm to 19.5 cm) and for home use.

When the device detects the appearance of irregular heartbeats during measurement, an indicated symbol will appear with measuring readings. And this device can let the memory data be transferred to the connected personal computer (PC) via USB cable.

Besides, Risk Category Indicator (Six Levels) feature will judge blood pressure results into six levels based on WHO (World Health Organization) classification with corresponding bar segment on the edge of screen.

Prescription Use _____
(Part 21 CFR 801 Subpart D)


AND/OR

Over-The-Counter Use V
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE
OF NEEDED)

Concurrence of CDRH, Office of Devices Evaluation (ODE)

Page 1 of 1



Owen P. Faris -5
2013.06.19 16:03:41
-04'00'